

REPUBLIQUE DU CAMEROUN
Paix – Travail – Patrie

MINISTRE DE LA SANTE PUBLIQUE

HOPITAL GENERAL DE YAOUNDE

DIRECTION GENERALE

DIRECTION DES AFFAIRES
ADMINISTRATIVES ET FINANCIERS

STRUCTURE INTERNE DE GESTION
ADMINISTRATIVE DES MARCHES
PUBLICS

B.P 5408 Yaoundé
Tél : (237) 222.21.31.81
Fax : (237) 222.21.20.15
hopitalgeneraldeyaounde@gmail.com



REPUBLIC OF CAMEROON
Peace – Work – Fatherland

MINISTRY OF PUBLIC HEALTH

GENERAL HOSPITAL OF YAOUNDE

GENERAL DIRECTORATE

DIRECTORATE OF ADMINISTRATIVE AND
FINANCIAL AFFAIRS

INTERNAL STRUCTURE FOR THE
ADMINISTRATIVE
MANAGEMENT OF PUBLIC
PROCUREMENT

P.O BOX: 5408 Yaoundé
hopitalgeneraldeyaounde@gmail.com

ADDITIF I

N° 207 7 026

/ A/HGY/ CIPM/SIGAMP/DU

09 JUL 2026

**AVIS D'APPEL D'OFFRES NATIONAL OUVERT N°09/AONO/HGY/CIPM/26
DU 11 JUIN 2026 EN PROCEDURE D'URGENCE POUR LA FOURNITURE ET L'INSTALLATION D'EQUIPEMENTS
MEDICAUX DE RADIOLOGIE MEDICALE AU SERVICE D'IMAGERIE
FINANCEMENT : Fonds propres-Exercice 2026-Ligne 24 44 04**

Le Directeur Général de l'Hôpital Général de Yaoundé, Autorité contractante, invite les entreprises ayant déjà acquis le Dossier de Consultation susmentionné, à bien vouloir se présenter au Secrétariat du Directeur des Affaires Administratives et Financières de l'Hôpital Général de Yaoundé, afin de retirer les nouvelles clauses à considérer.

1. CRITERES D'EVALUATION

1.1 Critères éliminatoires

- a. Absence du cautionnement de soumission accompagnée du récépissé CDEC ; Dans le cas de la soumission en ligne, le récépissé et la copie de de la caution de soumission, délivrés par la CDEC, sont jointes dans l'offre du soumissionnaire, tandis que leurs originaux sont remis à la Commission de passation des marchés compétentes, lors de la séance d'ouverture des offres.
- b. La non -production au-delà du délai de 48h après l'ouverture des plis d'une pièce du dossier administratif jugée non conforme ou absente lors de l'ouverture des plis ;
- c. Les fausses déclarations, manœuvres frauduleuses ou falsification des pièces
- d. Non satisfaction à 70% des critères essentiels ;
- e. Absence d'une capacité financière équivalent à 30% du montant prévisionnel du marché
- f. Non-respect de la composition de l'appareil de radiologie télécommandée avec fluoroscopie (Composition : 01 générateur haute tension, 01 tube à rayon X, 01 collimateur électrique, 01 console de commande, 01 détecteur à capteur plan, 01 table mobile pour patient, 01 poste travail, 02 vitres plombés).
- g. Non -respect des domaines d'application pour l'appareil de radiologie télécommandé avec fluoroscopie (Imagerie digestive et biliaire, Urologie (IVP, cystographie), Radiologie interventionnelle, Orthopédie post-opératoire, Examens en position spéciale (bassin, colonne, etc.),
- h. Non-respect de la conformité du générateur haute fréquence pour l'appareil de radiologie télécommandé avec fluoroscopie (Puissance de sortie : 65 KW minimum, Courant : 10 – 800 mA, Tension : 40 – 150 KV, Temps d'exposition : 0,1–800 ms, Alimentation électrique : 380 VAC, 50 Hz, Protection anti-surchauffe : Alarmes après 5 min, arrêt après 10 min),
- i. Non- respect de la conformité du tube à rayon X pour l'appareil de radiologie télécommandé avec fluoroscopie (Foyers : Petite foyer : 0,6 mm minimum, Grand foyer : 1,2 mm minimum, Type d'anode : Tournante, Capacité thermique de l'anode : 350 kHU minimum, Puissance d'entrée : Grand foyer : 70 KW minimum, Petit foyer : 30 KW minimum, Tension d'exposition : 40 - 150 KV),

- j. Non-respect de la conformité de la table mobile pour patient pour l'appareil de radiologie télécommandé avec fluoroscopie (Type : Table pour imagerie gastro-intestinale, Déplacement latéral du plateau : 250 mm minimum, Plage de rotation : -30° à +30° minimum, Hauteur de la surface de la table : 950 mm minimum, Capacité de charge : 130 kg minimum, Plage et déviation d'inclinaison de la table : +90° à 0° à -20°, Pédale de commande, Dimensions du plateau 2100 × 880 × 10 mm minimum, Structure en fibre de carbone, Panneau de commande supplémentaire sur le côté frontal de la table),
- k. Absence de prospectus, catalogue, dessin ou fiche technique produit par le fabricant ;
- l. Absence de l'agrément et du certificat de bonnes pratiques du MINSANTE
- m. Absence de l'autorisation délivrée par le fabricant
- n. Absence d'un prix unitaire quantifié dans l'Offre financière ;
- o. Absence d'une déclaration sur l'honneur de n'avoir pas abandonné de chantier durant les trois (03) dernières années
- p. Absence de la lettre de soumission.
- q. Absence d'une attestation capacité financière à hauteur de 48 000 000(quarante-huit millions) délivrée par une banque agréée
- r. Absence de la charte d'intégrité
- s. Absence de la déclaration d'engagement social et environnemental
- t. Non-respect du format de fichiers des offres
- u. Absence du certificat de conformité du fabricant (ISO, CE, IQNET...),
- v. Absence de l'attestation de visite de site signée sur l'honneur par le prestataire.

1.2 CRITERES ESSENTIELS

L'évaluation des offres technique sera faite suivant le système binaire (oui/non) sur la base des critères essentiels de qualification ci-dessous :

Les critères essentiels à la qualification des fournisseurs porteront sur :

- La présentation de l'offre ;
- Propositions techniques ;
- Les références du soumissionnaire ;
- Le délai de garantie ;
- Le calendrier de livraison ;
- La preuve d'acceptation des conditions du marché ;
- Le service après-vente (disponibilité des pièces de rechange) ;
- La qualification et l'expérience du personnel.

Seuls les soumissionnaires qui auront obtenu au moins 70% OUI seront jugés techniquement qualifiés pour ce lot et admis à l'analyse financière

2. OUVERTURE DES PLIS

Au lieu de : l'ouverture des offres (administratives, techniques et financières) aura lieu le 10 juillet 2026 à 14 heures, heure locale,

LIRE : l'ouverture des offres (administratives, techniques et financières) aura lieu le 24 Juillet 2026 à 14 heures, heure locale.

LE RESTE, SANS CHANGEMENT.

Yaoundé le

09 JUL 2026

LE DIRECTEUR GENERAL



2

P. Emmanuel Noel Emmanuel

REPUBLIQUE DU CAMEROUN
Paix – Travail – Patrie

MINISTRE DE LA SANTE PUBLIQUE

HOPITAL GENERAL DE YAOUNDE

DIRECTION GENERALE

DIRECTION DES AFFAIRES
ADMINISTRATIVES ET FINANCIERS


STRUCTURE INTERNE DE GESTION
ADMINISTRATIVE DES MARCHES
PUBLICS

B.P 5408 Yaoundé

Tél : (237) 222.21.31.81

Fax : (237) 222.21.20.15

hospitalgeneraldeyaounde@gmail.com



REPUBLIC OF CAMEROON
Peace – Work – Fatherland

MINISTRY OF PUBLIC HEALTH

GENERAL HOSPITAL OF YAOUNDE

GENERAL DIRECTORATE

DIRECTORATE OF ADMINISTRATIVE AND
FINANCIAL AFFAIRS

INTERNAL STRUCTURE FOR THE
ADMINISTRATIVE
MANAGEMENT OF PUBLIC
PROCUREMENT

P.O BOX: 5408 Yaoundé

hospitalgeneraldeyaounde@gmail.com

ADDENDUM I

No. **2027/Q26**

/ A/HGY/ CIPM/SIGAMP/DU

09 JUL 2026

**NATIONAL OPEN TENDER NO. 09/AONO/HGY/CIPM/26 OF JUNE, 11 2026, UNDER EMERGENCY PROCEDURE
FOR THE SUPPLY AND INSTALLATION OF MEDICAL RADIOLOGY EQUIPMENT FOR THE IMAGING
DEPARTMENT OF THE YAOUNDE GENERAL HOSPITAL
FINANCING: Equity - Fiscal Year 2026 - Line 24 44 04.**

The Director General of Yaoundé General Hospital, the Contracting Authority, invites companies that have already obtained the aforementioned Tender Documents to present themselves at the Secretariat of the Director of Administrative and Financial Affairs of Yaoundé General Hospital to collect the new clauses to be considered.

1. Evaluation Criteria

1.1. Elimination Criteria

- a. Absence of the bid security accompanied by the CDEC receipt; In the case of online submissions, the receipt and a copy of the bid security, issued by the CDEC, are included in the bidder's offer, while the originals are submitted to the competent Procurement Commission during the bid opening session.
- b. Failure to submit, within 48 hours of the bid opening, a document from the administrative file deemed non-compliant or missing at the time of the bid opening;
- c. False statements, fraudulent maneuvers, or falsification of documents;
- d. Failure to meet 70% of the essential criteria;
- e. Lack of financial capacity equivalent to 30% of the estimated market amount
- f. Failure to comply with the specifications of the remote-controlled fluoroscopy radiology unit (Composition: 1 high-voltage generator, 1 X-ray tube, 1 electric collimator, 1 control console, 1 flat-panel detector, 1 mobile patient table, 1 workstation, 2 lead-lined glass panes).
- g. Failure to comply with the application areas for the remote-controlled X-ray unit with fluoroscopy (Digestive and biliary imaging, Urology (IVP, cystography), Interventional radiology, Post-operative orthopedics, Examinations in special positions (pelvis, spine, etc.)),
- h. Failure to comply with the high-frequency generator specifications for the remote-controlled X-ray unit with fluoroscopy (Output power: 65 kW minimum, Current: 10–800 mA, Voltage: 40–150 kV, Exposure time: 0.1–800 ms, Power supply: 380 VAC, 50 Hz, Overheating protection: Alarms after 5 min, shutdown after 10 min).
- i. Non-compliance of the X-ray tube for the remote-controlled radiology device with fluoroscopy (Spots: Small spot: 0.6 mm minimum, Large spot: 1.2 mm minimum, Anode type: Rotating, Anode thermal capacity: 350 kHU minimum, Input power: Large spot: 70 KW minimum, Small spot: 30 KW minimum, Exposure voltage: 40 - 150 KV),
- j. Non-compliance of the mobile patient table for the remote-controlled X-ray unit with fluoroscopy (Type: Table for gastrointestinal imaging, Lateral movement of the tabletop: 250 mm minimum, Rotation range: -30° to +30° minimum, Table surface height: 950 mm minimum, Load capacity: 130 kg minimum, Table tilt range and deviation: +90° to 0° to -20°, Foot pedal, Tabletop dimensions: 2100 × 880 × 10 mm minimum, Carbon fiber structure, Additional control panel on the front side of the table),

- k. Absence of brochure, catalog, drawing, or technical data sheet produced by the manufacturer;
- l. Absence of approval and certificate of good practice from the Ministry of Health;
- m. Absence of authorization issued by the manufacturer;
- n. Absence of a quantified unit price in the Financial Offer;
- o. Absence of a sworn statement that no projects have been abandoned during the last three (3) years;
- p. Absence of the letter of submission;
- q. Absence of the integrity charter;
- r. Absence of the social and environmental commitment statement;
- s. Non-compliance with the bid file format;
- t. Absence of the manufacturer's certificate of conformity (ISO, CE, IQNET, etc.);
- v. Absence of the site visit certificate signed by the service provider.

1.2. Essential Criteria

The technical bids will be evaluated using a binary system (yes/no) based on the following essential qualification criteria:

- Presentation of the offer;
- Bidder's references;
- Warranty period;
- Delivery schedule;
- Proof of acceptance of contract terms;
- After-sales service (availability of spare parts, technical personnel);
- Access to a line of credit or other financial resources
- Qualification and experience of personnel

Only bidders who obtain at least 70% "YES" will be deemed technically qualified for this lot and will be admitted to the financial analysis.

2. OPENING OF BIDS

Instead of: The opening of bids (administrative, technical, and financial) will take place on July 10, 2026, at 2:00 p.m. local time.

READ: The opening of bids (administrative, technical, and financial) will take place on July 24, 2026, at 2:00 p.m. local time.

THE REST REMAINS UNCHANGED.

Yaoundé,

09 JUL 2026

THE DIRECTOR GENERAL



Dr. Emmanuël Noël Emmanuël

REPUBLIQUE DU CAMEROUN
Paix – Travail – Patrie

MINISTERE DE LA SANTE PUBLIQUE

HOPITAL GENERAL DE YAOUNDE

DIRECTION GENERALE

DIRECTION DES AFFAIRES
ADMINISTRATIVES ET FINANCIERS

STRUCTURE INTERNE DE GESTION
ADMINISTRATIVE DES MARCHES
PUBLICS

B.P 5408 Yaoundé
Tél : (237) 222.21.31.81
Fax : (237) 222.21.20.15
hospitalgeneraldeyaounde@gmail.com



REPUBLIC OF CAMEROON
Peace – Work – Fatherland

MINISTRY OF PUBLIC HEALTH

GENERAL HOSPITAL OF YAOUNDE

GENERAL DIRECTORATE

DIRECTORATE OF ADMINISTRATIVE AND
FINANCIAL AFFAIRS

INTERNAL STRUCTURE FOR THE
ADMINISTRATIVE
MANAGEMENT OF PUBLIC
PROCUREMENT

P.O BOX: 5408 Yaoundé
hospitalgeneraldeyaounde@gmail.com

ADDENDUM I

No. 2027 / Q26 / A/HGY/ CIPM/SIGAMP/DU 09 JUL 2026

NATIONAL OPEN TENDER NO. 09/AONO/HGY/CIPM/26 OF JUNE, 11 2026, UNDER EMERGENCY PROCEDURE FOR THE SUPPLY AND INSTALLATION OF MEDICAL RADIOLOGY EQUIPMENT FOR THE IMAGING DEPARTMENT OF THE YAOUNDE GENERAL HOSPITAL
FINANCING: Equity - Fiscal Year 2026 - Line 24 44 04.

The Director General of Yaoundé General Hospital, the Contracting Authority, invites companies that have already obtained the aforementioned Tender Documents to present themselves at the Secretariat of the Director of Administrative and Financial Affairs of Yaoundé General Hospital to collect the new clauses to be considered.

1. Evaluation Criteria

1.1. Elimination Criteria

- a. Absence of the bid security accompanied by the CDEC receipt; In the case of online submissions, the receipt and a copy of the bid security, issued by the CDEC, are included in the bidder's offer, while the originals are submitted to the competent Procurement Commission during the bid opening session.
- b. Failure to submit, within 48 hours of the bid opening, a document from the administrative file deemed non-compliant or missing at the time of the bid opening;
- c. False statements, fraudulent maneuvers, or falsification of documents;
- d. Failure to meet 70% of the essential criteria;
- e. Lack of financial capacity equivalent to 30% of the estimated market amount
- f. Failure to comply with the specifications of the remote-controlled fluoroscopy radiology unit (Composition: 1 high-voltage generator, 1 X-ray tube, 1 electric collimator, 1 control console, 1 flat-panel detector, 1 mobile patient table, 1 workstation, 2 lead-lined glass panes).
- g. Failure to comply with the application areas for the remote-controlled X-ray unit with fluoroscopy (Digestive and biliary imaging, Urology (IVP, cystography), Interventional radiology, Post-operative orthopedics, Examinations in special positions (pelvis, spine, etc.)),
- h. Failure to comply with the high-frequency generator specifications for the remote-controlled X-ray unit with fluoroscopy (Output power: 65 kW minimum, Current: 10–800 mA, Voltage: 40–150 kV, Exposure time: 0.1–800 ms, Power supply: 380 VAC, 50 Hz, Overheating protection: Alarms after 5 min, shutdown after 10 min).
- i. Non-compliance of the X-ray tube for the remote-controlled radiology device with fluoroscopy (Spots: Small spot: 0.6 mm minimum, Large spot: 1.2 mm minimum, Anode type: Rotating, Anode thermal capacity: 350 kHU minimum, Input power: Large spot: 70 KW minimum, Small spot: 30 KW minimum, Exposure voltage: 40 - 150 KV),
- j. Non-compliance of the mobile patient table for the remote-controlled X-ray unit with fluoroscopy (Type: Table for gastrointestinal imaging, Lateral movement of the tabletop: 250 mm minimum, Rotation range: -30° to +30° minimum, Table surface height: 950 mm minimum, Load capacity: 130 kg minimum, Table tilt range and deviation: +90° to 0° to -20°, Foot pedal, Tabletop dimensions: 2100 × 880 × 10 mm minimum, Carbon fiber structure, Additional control panel on the front side of the table),

- k. Absence of brochure, catalog, drawing, or technical data sheet produced by the manufacturer;
- l. Absence of approval and certificate of good practice from the Ministry of Health;
- m. Absence of authorization issued by the manufacturer;
- n. Absence of a quantified unit price in the Financial Offer;
- o. Absence of a sworn statement that no projects have been abandoned during the last three (3) years;
- p. Absence of the letter of submission;
- q. Absence of the integrity charter;
- r. Absence of the social and environmental commitment statement;
- s. Non-compliance with the bid file format;
- t. Absence of the manufacturer's certificate of conformity (ISO, CE, IQNET, etc.);
- v. Absence of the site visit certificate signed by the service provider.

1.2. Essential Criteria

The technical bids will be evaluated using a binary system (yes/no) based on the following essential qualification criteria:

- Presentation of the offer;
- Bidder's references;
- Warranty period;
- Delivery schedule;
- Proof of acceptance of contract terms;
- After-sales service (availability of spare parts, technical personnel);
- Access to a line of credit or other financial resources
- Qualification and experience of personnel

Only bidders who obtain at least 70% "YES" will be deemed technically qualified for this lot and will be admitted to the financial analysis.

2. OPENING OF BIDS

Instead of: The opening of bids (administrative, technical, and financial) will take place on July 10, 2026, at 2:00 p.m. local time.

READ: The opening of bids (administrative, technical, and financial) will take place on July 24, 2026, at 2:00 p.m. local time.

THE REST REMAINS UNCHANGED.

Yaoundé,

09 JUL 2026

THE DIRECTOR GENERAL



Dr. Emmanuël Noël Emmanuël